

Lead Investigator

Wenqing (Vicky) Xu PhD, Professor, Director of the Center for Human-Environmental Systems, College of Engineering, Department of Civil and Environmental Engineering.

Background and Unmet Need

Per- and poly-fluoroalkyl (PFAS), also known as 'Forever Chemicals,' this class of chemicals has been a growing contaminant of concern throughout the last decade. PFAS in the US Environmental Protection Agency's (EPA) list of priority pollutants. PFAS in drinking water, PFOA (4ppt)

⁴, PFOS (4 ppt), PFHxS (10ppt), PFNA (10ppt), HFPO-DA (10 ppt), and PFAS mixtures containing at least two or more of PFHxS, PFNA, HFPO-DA, and PFBS (hazard index of 1).⁵ These new standards have fueled a strong demand for developing novel technologies to quantify and to identify PFAS at or below regulatory standards.

The current standard for detecting PFAS are liquid chromatography-triple quadrupole-tandem mass spectrometry (LC-MS/MS). This technology is expensive to obtain and operate, requires specialized operator training, and is labor and time intensive. There is a growing consensus among the scientific community that technologies for rapid PFAS detection at and below the regulatory concentration are critical but lacking.

Experts believe the demand for rapid, mobile, and economical PFAS detection will continue to grow as global governmental agencies evolve regulatory requirements on monitoring PFAS contamination.

Opportunity

Dr. Xu has developed a novel Molecularly Imprinted Polymer (MIP)-enabled electrochemical platform for rapid and direct quantification of non-electroactive PFAS. Multiple PFAS's within water samples simultaneously can be detected using this platform. The detection process has been demonstrated in laboratory-scale and field reusability test.

¹ *Our Current Understanding of the Human Health and Environmental Risks of PFAS.*, EPA, n.d.

² *Per- and polyfluoroalkyl substances (PFAS) in United States tap water: Comparison of underserved private-well and public-supply exposures and associated health implications.*, Elsevier, June 2023.

³ *Population-Wide Exposure to Per- and Polyfluoroalkyl Substances from Drinking Water in the United States.*, ACS Publications, October 2020.

⁴

This proprietary MIP technology is a platform for future PFAS sensors. The technology may be integrated into a compact mobile electrochemical platform, enabling on-site sample collection and detection. Compared to the current state of the art, on-site detection of PFAS contaminants is not technically feasible. As a result, the technology platform presents the opportunity to be further developed into a first-of-its-kind field-deployable sensor to provide real-time detection of PFAS compounds while in the field.

The PFAS testing market was valued at \$110 million in 2023 and is expected to grow to \$217 million by 2028 at a CAGR of 14.4%. The market's growth is driven by the privatization of environmental testing services, increased investment in water resource management, and regulatory pressure from the Environmental Protection Agency.⁷

Unique Attributes

- Rapid detection and quantification of PFAS at sub ng L-1 concentrations in water
- Platform technology compatible with the development of mobile sensor apparatus.
- Economical solution to PFAS detection in varying environments.
- Direct quantification of PFAS enabled by a novel and redox-active Molecular Imprinted Polymer platform that eliminates the need for external redox probes.

Applications

A redox-active molecular imprinted polymer for the rapid ex-situ detection of PFAS contaminants within drinking water resources.

Stage of Development

Technology validated in a relevant environment (Technology Readiness Level 5) and operational prototype.